



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

CALIBRATOR FOR ACETATES IN URINE

(ACRYLONITRILE, BUTHYLACETATE, HEPTANE, ETHYLACETATE, ISOBUTYLACETATE, TETRAHYDROFURANE)

Code GC30016
TO USE WITH KIT CODE GC30010

OBJECTIVE


These calibrators are used to calibrate the GC/MS system in the quantitative determination of the analyte contained in them. Each analyte has a concentration such that a 5-points calibration curve can be constructed. These calibrators are in human matrix and should be handled as if they were a real patient sample.

STORAGE AND STABILITY

The calibrators are stable 36 months from the date of preparation if stored at -25/-15°C. Do not use them after the expiry date.

PRECAUTIONS

These calibrators in human matrix must be treated with care and considered as potentially infectious.

LOT		


AVAILABLE PACKAGINGS:

- GC30016 CALIBRATOR FOR ACETATES IN URINE

7 x 2 x 10 ml

N° 004	06/2024
--------	---------

CONCENTRATIONS
CODE GC30016

LOT		

ANALYTE	MEASUREMENT UNIT	C0	C1	C2	C3	C4
ACRYLONITRILE	mg/l					
BUTHYLACETATE	mg/l					
HEPTANE	mg/l					
ETHYLACETATE	mg/l					
ISOBUTHYLACETATE	mg/l					
TETRAHYDROFURANE	mg/l					

This product fulfils all the requirements of Directive 98/79/EC of 27/10/1998 on *in vitro* diagnostic medical devices (IVD). The declaration of conformity is available upon request.



FOR IN VITRO DIAGNOSTIC USE ONLY



EUREKA S.R.L. LAB DIVISION
Via M. D'Antona 28
60033 Chiaravalle (AN) - Italy
Tel +39 071 7450790
eureka-support@sentinel.it
www.eurekakit.com