



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

CALIBRATOR FOR VOLATILE ORGANIC COMPOUNDS IN WHOLE BLOOD

(ACETONE, BENZENE, DICHLOROMETHANE, ETHANOL, METHANOL, PERCHLOROETHYLENE, TOLUENE)

Code GC77116
TO USE WITH KIT CODE GC77110

OBJECTIVE



These calibrators are used to calibrate the GC/MS system in the quantitative determination of the analyte contained in them. Each analyte has a concentration such that a 7-point calibration curve can be constructed. These calibrators are in human matrix and should be handled as if they were a real patient sample.

STORAGE AND STABILITY

The calibrators are stable 36 months from the date of preparation if stored at -25/-15°C. Do not use them after the expiry date.

PRECAUTIONS

These calibrators in human matrix must be treated with care and considered as potentially infectious.

LOT		

AVAILABLE PACKAGINGS:



GC77116 CALIBRATOR FOR VOC IN WHOLE BLOOD

7 x 2 x 10 ml

N° 004

11/2024

CONCENTRATIONS
CODE GC77116

LOT		

ANALYTE	MEASUREMENT UNIT	C0	C1	C2	C3	C4	C5	C6
ACETONE	mg/l	0	1,0	5,0	10,0	25,0	50,0	100,0
BENZENE	mg/l	0	0,01	0,05	0,1	0,25	0,5	1,0
DICHLOROMETHANE	mg/l	0	0,01	0,05	0,1	0,25	0,5	1,0
ETHANOL	mg/l	0	1,0	5,0	10,0	25,0	50,0	100,0
METHANOL	mg/l	0	1,0	5,0	10,0	25,0	50,0	100,0
PERCHLOROETHYLENE	mg/l	0	0,01	0,05	0,1	0,25	0,5	1,0
TOLUENE	mg/l	0	0,01	0,05	0,1	0,25	0,5	1,0

This product fulfils all the requirements of Directive 98/79/EC of 27/10/1998 on *in vitro* diagnostic medical devices (IVD). The declaration of conformity is available upon request.

CE

FOR *IN VITRO* DIAGNOSTIC USE ONLY



EUREKA S.R.L. LAB DIVISION
Via M. D'Antona 28
60033 Chiaravalle (AN) - Italy
Tel +39 071 7450790
eureka-support@sentinel.it
www.eurekakit.com