



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

CONTROLS FOR VOLATILE ORGANIC COMPOUNDS IN WHOLE BLOOD - LEVELS 1 and 2

(ACETONE, BENZENE, DICHLOROMETHANE, ETHANOL, METHANOL, PERCHLOROETHYLENE,
TOLUENE)

Code GC77119

OBJECTIVE



These controls are used for internal quality control and serve to control the accuracy and precision of analytical procedures dedicated to the quantitative determination of the analytes they contain. These controls are in human matrix and are available in two different concentration ranges.

STORAGE AND STABILITY

The controls are stable 36 months from the date of preparation if stored at -25/-15°C. Do not use them after the expiry date.

PRECAUTIONS

These controls in human matrix must be treated with care and considered as potentially infectious.

LOT		



AVAILABLE PACKAGING:

- GC77119 CONTROL FOR VOC IN WHOLE BLOOD - LEVELS 1 and 2

2 x 10 x 10 ml

N° 004	11/2024
--------	---------

CONCENTRATIONS
CODE GC77119

LOT		

ANALYTE	MEASUREMENT UNIT	MEDIUM VALUE L1	RANGE L1	MEDIUM VALUE L2	RANGE L2
ACETONE	mg/l	5,0	3,5 – 6,5	50,0	35,0 – 65,0
BENZENE	mg/l	0,05	0,035 – 0,065	0,5	0,35 – 0,65
DICHLOROMETHANE	mg/l	0,05	0,035 – 0,065	0,5	0,35 – 0,65
ETHANOL	mg/l	5,0	3,5 – 6,5	50,0	35,0 – 65,0
METHANOL	mg/l	5,0	3,5 – 6,5	50,0	35,0 – 65,0
PERCHLOROETHYLENE	mg/l	0,05	0,035 – 0,065	0,5	0,35 – 0,65
TOLUENE	mg/l	0,05	0,035 – 0,065	0,5	0,35 – 0,65

This product fulfils all the requirements of Directive 98/79/EC of 27/10/1998 on *in vitro* diagnostic medical devices (IVD). The declaration of conformity is available upon request.



FOR IN VITRO DIAGNOSTIC USE ONLY



EUREKA S.R.L. LAB DIVISION
Via M. D'Antona 28
60033 Chiaravalle (AN) - Italy
Tel +39 071 7450790
eureka-support@sentinel.it
www.eurekakit.com