



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

CONTROLS FOR VOLATILE ORGANIC COMPOUNDS IN WHOLE BLOOD - LEVELS 1 and 2

(ACETONE, BENZENE, DICHLOROMETHANE, ETHANOL, METHANOL, PERCHLOROETHYLENE,
TOLUENE)

Code GC77119

OBJECTIVE




These controls are used for internal quality control and serve to control the accuracy and precision of analytical procedures dedicated to the quantitative determination of the analytes they contain. These controls are in human matrix and are available in two different concentration ranges.

STORAGE AND STABILITY

The controls are stable 36 months from the date of preparation if stored at -25/-15°C. Do not use them after the expiry date.

PRECAUTIONS

These controls in human matrix must be treated with care and considered as potentially infectious.

		
001	March 2022	March 2025

AVAILABLE PACKAGING:




- GC77119 CONTROL FOR VOC IN WHOLE BLOOD - LEVELS 1 and 2

2 x 10 x 10 ml

Release N° 003

May 2019

CONCENTRATIONS
CODE GC77119

 LOT		
001	March 2022	March 2025

<i>ANALYTE</i>	<i>MEASUREMENT UNIT</i>	<i>MEDIUM VALUE L1</i>	<i>RANGE L1</i>	<i>MEDIUM VALUE L2</i>	<i>RANGE L2</i>
ACETONE	mg/l	5,0	3,5 – 6,5	50,0	35,0 – 65,0
BENZENE	mg/l	0,05	0,035 – 0,065	0,5	0,35 – 0,65
DICHLOROMETHANE	mg/l	0,05	0,035 – 0,065	0,5	0,35 – 0,65
ETHANOL	mg/l	5,0	3,5 – 6,5	50,0	35,0 – 65,0
METHANOL	mg/l	5,0	3,5 – 6,5	50,0	35,0 – 65,0
PERCHLOROETHYLENE	mg/l	0,05	0,035 – 0,065	0,5	0,35 – 0,65
TOLUENE	mg/l	0,05	0,035 – 0,065	0,5	0,35 – 0,65

This product fulfils all the requirements of Directive 98/79/EC of 27/10/1998 on *in vitro* diagnostic medical devices (IVD). The declaration of conformity is available upon request.



FOR *IN VITRO* DIAGNOSTIC USE ONLY

produced by

COMPANY WITH
MANAGEMENT SYSTEM
CERTIFIED BY DNV
= ISO 9001 =
= ISO 13485 =



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

EUREKA srl – LAB DIVISION ,
Via Enrico Fermi 25, 60033 Chiaravalle (AN) ITALY
Tel. + 39 071 74 50 790
Fax. + 39 071 74 96 579
info@eurekaone.com
www.eurekakit.com



Release N° 003

May 2019