



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS  
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

## CONTROL LYOPHIL. IN URINE FOR DRUGS OF ABUSE – LEVELS 1 and 2 ( $\Delta^9$ -THC-COOH)

Code CC47019

### OBJECTIVE

These controls in urine are used for internal quality control and serve to keep under control the accuracy and precision of the analytical procedures dedicated to the quantitative determination of the analytes contained therein. These controls are lyophilized in human matrix and are available in two different concentration range.

### RECONSTITUTION



Remove the metal seal and rubber stopper from the vial. Add exactly 10 ml of HPLC grade H<sub>2</sub>O in the vial. Replace the rubber stopper, shake and let stand for 5 to 10 minutes. Before use, mix by inverting the vial to dissolve the material until a homogeneous clear solution.

### STORAGE AND STABILITY

The controls are stable for 36 months from the date of preparation if stored at 2-8 ° C. After reconstitution are stable 2 days at 2-8 ° C and 2 months at -20 ° C. Do not use after the expiry date.

### PRECAUTIONS

These controls in human matrix should be treated with care and treated as potentially infectious.

LOT		


### PACKAGING AVAILABLE:

- CC47019 CONTROL IN URINE FOR DRUGS OF ABUSE - LEVELS 1 and 2    2 x 6 x 10 ml

N° 004

06/2024

CONCENTRATIONS  
CODE CC47019

LOT		

ANALYTE	MEASUREMENT UNIT	L1	RANGE	MEDIUM VALUE OBTAINED + DS*	L2	RANGE	MEDIUM VALUE OBTAINED + DS*
$\Delta^9$ -THC-COOH	ng/ml						

\* The procedures for production and validation / quality control of the individual lot shall be performed using the chromatographic method LC / MS / MS validated according to international guidelines of the Food and Drug Administration (FDA-May 2001) and the International Conference of Harmonisation of technical requirements for registration of pharmaceuticals for human use (ICH-Q2-R1-Nov 2005). The mean value and standard deviation were obtained in the validation phase of batch processing for level four bottles in duplicate.

This product fulfils all the requirements of Directive 98/79/EC of 27/10/1998 on *in vitro* diagnostic medical devices (IVD). The declaration of conformity is available upon request.



**FOR IN VITRO DIAGNOSTIC USE ONLY**



EUREKA S.R.L. LAB DIVISION  
Via M. D'Antona 28  
60033 Chiaravalle (AN) - Italy  
Tel +39 071 7450790  
eureka-support@sentinel.it  
www.eurekakit.com