



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

CALIBRATOR LYOPHIL. IN PLASMA FOR ANTIMYCOTICS
(5-Flucytosine, Anidulafungin, Caspofungin, Fluconazole, Ketoconazole, OH-Itraconazole, Isovucanazole, Itraconazole, Miconazole, Posaconazole, Voriconazole)

Code LC82016
TO USE WITH KIT CODE LC82010

OBJECTIVE

The calibrators in plasma are used to calibrate the HPLC system for the quantitative measurement of analytes contained therein. Each analyte has a concentration that allows the construction of a calibration curve of seven points. The calibrators are freeze-dried human matrix and should be handled as if they were a real patient sample.

RECONSTITUTION

Remove the metal seal and rubber stopper from the vial. Add exactly 1 ml of HPLC grade H₂O in the vial. Replace the rubber stopper, shake and let stand for 5 to 10 minutes. Before use, mix by inverting the vial to dissolve the material until a homogeneous clear solution.

STORAGE AND STABILITY

The calibrators are stable for 36 months from the date of preparation if stored at +2 °C.

After reconstitution are stable 7 days at +2 °C and 1 month at -25 °C. Do not use after the expiry date.

PRECAUTIONS

These calibrators in human matrix should be treated with care and treated as potentially infectious.

LOT		

PACKAGING AVAILABLE:

- LC82016 CALIBRATOR IN PLASMA FOR ANTIMYCOTICS 7 x 2 x 1 ml

N° 003	06/2024
--------	---------

CONCENTRATIONS
CODE LC82016

LOT		

ANALYTE	MEASUREMENT UNIT	C0	C1	C2	C3	C4	C5	C6
5-FLUCYTOSINE	mg/l	N.D.						
ANIDULAFUNGIN	mg/l	N.D.						
CASPOFUNGIN	mg/l	N.D.						
FLUCONAZOLE	mg/l	N.D.						
KETOCONAZOLE	mg/l	N.D.						
OH-ITRACONAZOLE	mg/l	N.D.						
ISAVUCONAZOLE	mg/l	N.D.						
ITRACONAZOLE	mg/l	N.D.						
MICAFUNGIN	mg/l	N.D.						
POSACONAZOLE	mg/l	N.D.						
VORICONAZOLE	mg/l	N.D.						

This product fulfills all the requirements of Directive 98/79/EC of 27/10/1998 on in vitro diagnostic medical devices (IVD). The declaration of conformity is available upon request.

CE

FOR IN VITRO DIAGNOSTIC USE ONLY



EUREKA S.R.L. LAB DIVISION
Via M. D'Antona 28
60033 Chiaravalle (AN) - Italy
Tel +39 071 7450790
eureka-support@sentinel.it
www.eurekakit.com