



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

CALIBRATOR IN PLASMA FOR PONATINIB

Code LC99016
TO USE WITH KIT CODE LC99010

OBJECTIVE

The calibrators in plasma are used to calibrate the HPLC system for the quantitative measurement of analytes contained therein. Each analyte has a concentration that allows the construction of a calibration curve of six points. The calibrators are freeze-dried human matrix and should be handled as if they were a real patient sample.

RECOVERY

Remove the metal seal and rubber stopper from the vial. Add exactly 1 ml of HPLC-grade water through the tube. Replace the rubber stopper, allow to stand for at least 5 to 10 minutes. Before use, mix by inverting the vial to dissolve the material until a homogeneous solution.

STORAGE AND STABILITY

The calibrators are stable for 36 months from the date of preparation if stored at 2-8 ° C. After reconstitution are stable 1 week at 2-8 °C and 1 month at -20 ° C. Do not use after the expiry date.

PRECAUTIONS

These calibrators in human matrix should be treated with care and treated as potentially infectious.

LOT		

PACKAGING AVAILABLE:

- LC99016 CALIBRATOR IN PLASMA FOR PONATINIB 7 x 2 x 1 ml

N° 003	10/2024
--------	---------

CONCENTRATIONS
CODE LC99016

LOT		

ANALYTE	MEASUREMENT UNIT	C0	C1	C2	C3	C4	C5	C6
PONATINIB	ng/ml							

This product fulfils all the requirements of Directive 98/79/EC of 27/10/1998 on *in vitro* diagnostic medical devices (IVD). The declaration of conformity is available upon request.



FOR IN VITRO DIAGNOSTIC USE ONLY



EUREKA S.R.L. LAB DIVISION
Via M. D'Antona 28
60033 Chiaravalle (AN) - Italy
Tel +39 071 7450790
eureka-support@sentinel.it
www.eurekakit.com