AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

CALIBRATOR FOR MYCOPHENOLIC ACID IN PLASMA

Code LC75316

TO USE WITH KIT CODE LC75310-15, LC75410

OBJECTIVE

The calibrators in plasma are used to calibrate the HPLC system for the quantitative measurement of analytes contained therein. Each analyte has a concentration that allows the construction of a calibration curve of seven points. The calibrators are freeze-dried human matrix and should be handled as if they were a real patient sample.

RECOVERY

Remove the metal seal and rubber stopper from the vial. Add exactly 1 ml of HPLC-grade water through the tube. Replace the rubber stopper, allow to stand for at least 30 to 40 minutes. Before use, mix by inverting the vial to dissolve the material until a homogeneous solution.

STORAGE AND STABILITY

The calibrators are stable for 36 months from the date of preparation if stored at -20 ° C. After reconstitution are stable 1 week at 2-8 °C and 1 month at -20 ° C. Do not use after the expiry date.

PRECAUTIONS

These calibrators in human matrix should be treated with care and treated as potentially infectious.

LOT		

PACKAGING AVAILABLE:

- LC75316 CALIBRATOR FOR MYCOPHENOLIC ACID IN PLASMA ***7 x 2 x 1 ml***

N° 005	**11/2025**

CONCENTRATION
CODE LC75316

LOT		

ANALYTE	*MEASUREMENT UNIT*	*C0*	*C1*	*C2*	*C3*	*C4*	*C5*	*C6*
Mycophenolic Acid	**µg/ml**	**N.D.**						

This product fulfills all the requirements of Directive 98/79/EC of 27/10/1998 on in vitro diagnostic medical devices (IVD). The declaration of conformity is available upon request.

CE

FOR *IN VITRO* DIAGNOSTIC USE ONLY

EUREKA S.R.L. LAB DIVISION
Via M. D'Antona 28
60033 Chiaravalle (AN) - Italy
Tel +39 071 7450790
eureka-support@sentinel.it
www.eurekakit.com